### § 888.3650

prosthesis that was in commercial distribution before May 28, 1976. Any other shoulder joint metal/metal or metal/polymer constrained cemented prosthesis shall have an approved PMA or a declared completed PDP in effect before being placed in commercial distribution.

[52 FR 33702, Sept. 4, 1987, as amended at 61 FR 50711, Sept. 27, 1996]

#### §888.3650 Shoulder joint metal/polymer non-constrained cemented prosthesis.

- (a) Identification. A shoulder joint metal/polymer non-constrained mented prosthesis is a device intended to be implanted to replace a shoulder joint. The device limits minimally (less than normal anatomic constraints) translation in one or more planes. It has no linkage across-the-joint. This generic type of device includes prostheses that have a humeral component made of alloys, such as cobalt-chromium-molybdenum, and a glenoid resurfacing component made of ultrahigh molecular weight polyethylene, and is limited to those prostheses intended for use with bone cement (§888.3027).
  - (b) Classification. Class III.
- (c) Date PMA or notice of completion of a PDP is required. No effective date has been established of the requirement for premarket approval. See §888.3.

## § 888.3660 Shoulder joint metal/polymer semi-constrained cemented prosthesis.

- (a) Identification. A shoulder joint metal/polymer semi-constrained mented prosthesis is a device intended to be implanted to replace a shoulder joint. The device limits translation and rotation in one or more planes via the geometry of its articulating surfaces. It has no linkage across-the-joint. This generic type of device includes prostheses that have a humeral resurfacing component made of alloys, such as cobalt-chromium-molybdenum, glenoid resurfacing component made of ultra-high molecular weight polyethylene, and is limited to those prostheses intended for use with bone cement (§888.3027).
  - (b) Classification. Class III.

(c) Date PMA or notice of completion of a PDP is required. No effective date has been established of the requirement for premarket approval. See §888.3.

### § 888.3680 Shoulder joint glenoid (hemi-shoulder) metallic cemented prosthesis.

- (a) Identification. A shoulder joint glenoid (hemi-shoulder) metallic cemented prosthesis is a device that has a glenoid (socket) component made of alloys, such as cobalt-chromium-molybdenum, or alloys with ultra-high molecular weight polyethylene and intended to be implanted to replace part of a shoulder joint. This generic type of device is limited to those prostheses intended for use with bone cement (§888.3027).
  - (b) Classification. Class III.
- (c) Date PMA or notice of completion of a PDP is required. A PMA or a notice of completion of a PDP is required to be filed with the Food and Drug Administration on or before December 26, 1996 for any shoulder joint glenoid (hemishoulder) metallic cemented prosthesis that was in commercial distribution before May 28, 1976, or that has, on or before December 26, 1996 been found to be substantially equivalent to a shoulder joint glenoid (hemi-shoulder) metallic cemented prosthesis that was in commercial distribution before May 28, 1976. Any other shoulder joint glenoid (hemi-shoulder) metallic cemented prosthesis shall have an approved PMA or a declared completed PDP in effect before being placed in commercial distribution.

[52 FR 33702, Sept. 4, 1987, as amended at 61 FR 50711, Sept. 27, 1996]

# § 888.3690 Shoulder joint humeral (hemi-shoulder) metallic uncemented prosthesis.

- (a) Identification. A shoulder joint humeral (hemi-shoulder) metallic uncemented prosthesis is a device made of alloys, such as cobalt-chromium-molybdenum. It has an intramedullary stem and is intended to be implanted to replace the articular surface of the proximal end of the humerus and to be fixed without bone cement (§888.3027). This device is not intended for biological fixation.
  - (b) Classification. Class II.